



Variations in Respiratory Excretion of Carbon Dioxide Can Be Used to Calculate Pulmonary Blood Flow

Citation

Preiss, David A., Takafumi Azami, and Richard D. Urman. 2015. "Variations in Respiratory Excretion of Carbon Dioxide Can Be Used to Calculate Pulmonary Blood Flow." *Journal of Clinical Medicine Research* 7 (2): 83-90. doi:10.14740/jocmr1979w. <http://dx.doi.org/10.14740/jocmr1979w>.

Published Version

doi:10.14740/jocmr1979w

Permanent link

<http://nrs.harvard.edu/urn-3:HUL.InstRepos:13454808>

Terms of Use

This article was downloaded from Harvard University's DASH repository, and is made available under the terms and conditions applicable to Other Posted Material, as set forth at <http://nrs.harvard.edu/urn-3:HUL.InstRepos:dash.current.terms-of-use#LAA>

Share Your Story

The Harvard community has made this article openly available.
Please share how this access benefits you. [Submit a story](#).

[Accessibility](#)

Variations in Respiratory Excretion of Carbon Dioxide Can Be Used to Calculate Pulmonary Blood Flow

David A. Preiss^a, Takafumi Azami^b, Richard D. Urman^{c, d}

Abstract

Background: A non-invasive means of measuring pulmonary blood flow (PBF) would have numerous benefits in medicine. Traditionally, respiratory-based methods require breathing maneuvers, partial rebreathing, or foreign gas mixing because exhaled CO₂ volume on a per-breath basis does not accurately represent alveolar exchange of CO₂. We hypothesized that if the dilutional effect of the functional residual capacity was accounted for, the relationship between the calculated volume of CO₂ removed per breath and the alveolar partial pressure of CO₂ would be reversely linear.

Methods: A computer model was developed that uses variable tidal breathing to calculate CO₂ removal per breath at the level of the alveoli. We iterated estimates for functional residual capacity to create the best linear fit of alveolar CO₂ pressure and CO₂ elimination for 10 minutes of breathing and incorporated the volume of CO₂ elimination into the Fick equation to calculate PBF.

Results: The relationship between alveolar pressure of CO₂ and CO₂ elimination produced an $R^2 = 0.83$. The optimal functional residual capacity differed from the "actual" capacity by 0.25 L (8.3%). The repeatability coefficient leveled at 0.09 at 10 breaths and the difference between the PBF calculated by the model and the preset blood flow was 0.62 ± 0.53 L/minute.

Conclusions: With variations in tidal breathing, a linear relationship exists between alveolar CO₂ pressure and CO₂ elimination. Existing technology may be used to calculate CO₂ elimination during quiet breathing and might therefore be used to accurately calculate PBF in humans with healthy lungs.

Keywords: Pulmonary blood flow; Cardiac output; Carbon dioxide elimination; End-tidal carbon dioxide

Introduction

A clinician's ability to determine a patient's cardiac output is critical in the assessment of cardiopulmonary health. Clinically, this measurement can be required intermittently during surgery, as well as post-operatively. Cardiac output may also be measured in research settings, but doing so with invasive methods requires expensive critical care equipment, trained personnel, a sterile environment and readily available resuscitative equipment.

Cardiac output is arguably one of the most important parameters reflecting cardiovascular health and yet its measurement is currently limited to patients with Swan-Ganz catheters in the operating room, critical and intensive care units. To date, no non-invasive method has proven itself comparable to thermodilution in accuracy, repeatability, and in physiological soundness [1, 2]. Nevertheless, the thermodilution method is plagued with limitations in clinical practice, the most important ones being associated with its invasiveness [3].

Methods for measuring cardiac output non-invasively can be classified as respiratory-based or non-respiratory-based. The latter, which includes Doppler (external, transtracheal and transesophageal), bioimpedance, and pulse contour analysis, has not yet been accepted into clinical practice because of either theoretical or practical limitations [4]. Additionally, there are some respiratory-based methods that utilize foreign gas breathing, including argon and acetylene, but these methods require a source of external gas, are impractical in many settings, and will not be dealt with in detail here [5].

The respiratory-based methods are the oldest, the most physiologically sound, and traditionally the most accepted method of cardiac output measurement. Their principles and assumptions are well understood as are their limitations, the most important of which is that they more accurately measure pulmonary blood flow (PBF) rather than cardiac output. Originally described for oxygen, Fick equation, developed in the late 1800s, is based on the measurement of elements of the mass balance across the lungs [6]. The following is the

Manuscript accepted for publication October 01, 2014

^aDepartment of Anesthesia, Critical Care, and Pain Medicine, Harvard Medical School/Massachusetts General Hospital, Boston, MA, USA

^bDepartment of Pathophysiology and Anesthesia, Nagoya City University School of Nursing, Nagoya City, Japan

^cDepartment of Anesthesia, Perioperative and Pain Medicine, Harvard Medical School/Brigham and Women's Hospital, Boston, MA, USA

^dCorresponding Author: Richard D. Urman, Department of Anesthesiology, Critical Care and Pain Medicine, Brigham and Women's Hospital/Harvard Medical School, Boston, MA 02115, USA. Email: rurman@partners.org

doi: <http://dx.doi.org/10.14740/jocmr1979w>




Figure 1. The relationship between \dot{V}_{CO_2A} and $P_A CO_2$ is linear. The slope of this line is proportional to the pulmonary blood flow. An intersecting line can be drawn from the origin following the relationship $V_A = \dot{V}_{CO_2} \times P_A CO_2 / 713$. The steady state point (A) for $P_A CO_2$ and \dot{V}_{CO_2A} is produced by the balance of CO₂ diffusion into the alveoli and the flow of blood into the pulmonary capillaries.

equivalent equation describing the movement of carbon dioxide (CO₂):

$$\text{Eq. 1: } PBF = \dot{V}_{CO_2} / (C_v CO_2 - C_a CO_2)$$

where PBF is the pulmonary blood flow, \dot{V}_{CO_2} is the metabolic CO₂ production, and $C_v CO_2$ and $C_a CO_2$ are the concentrations of CO₂ in the blood entering (mixed venous, v) and leaving (arterial, a) the lungs respectively. To keep the method purely non-invasive, alveolar pressure is conventionally used as a surrogate for content. The relationship between $P_a CO_2$ and $C_a CO_2$ is usually assumed to be linear in the physiologic range, and through previous *in-vitro*, CO₂ content of oxygenated blood can be estimated using the following equation [7, 8]:

$$\text{Eq. 2: } \text{Content} = 4 \times \text{Pressure} + 260$$

Generally, $P_a CO_2$ is approximated from alveolar CO₂ pressure ($P_A CO_2$), which in turn is approximated from end-tidal CO₂ pressure ($P_{ET} CO_2$). Mixed venous pressure of CO₂ ($P_v CO_2$) however, is difficult to obtain non-invasively, and has classically posed the biggest challenge to measurement of PBF. Respiratory maneuvers such as breathholding and rapid equilibration with external reservoirs have been employed to estimate $P_v CO_2$, but these have the disadvantage of requiring an external supply of CO₂, and may pose a challenge to patients with respiratory compromise [9]. Single-breath methods have been presented, but none has proven itself repeatable and accurate.

In 1980, a novel method of calculating PBF without the need for measuring $C_v CO_2$ was described by Gedeon et al [9]. That method demonstrated that if a subject's alveolar ventilation (V_A) were acutely and transiently reduced, a step change in $P_A CO_2$ and \dot{V}_{CO_2} would result (once steady state is reached, after a few breaths). At this point, assuming PBF and $P_v CO_2$ had not changed in this short time (< 30 - 45 s), two Fick equations could be applied to the model:

$$\text{Eq. 3a: } PBF = \dot{V}_{CO_2} / (S \times P_v CO_2 - S \times P_A CO_2)$$

$$\text{Eq. 3b: } PBF = \dot{V}_{CO_2}' / (S \times P_v CO_2 - S \times P_A CO_2')$$

where \dot{V}_{CO_2}' and $P_A CO_2'$ are the exhaled CO₂ volumes and alveolar CO₂ pressures after a new steady state has oc-




Figure 2. Gedeon's method of reducing V_A for several breaths acutely increases $P_A CO_2$ and reduces \dot{V}_{CO_2A} creating a new steady state point on the same line. Point A and point B can be used together to calculate the slope of the line, which is the pulmonary blood flow.

curred, respectively, and S is a conversion constant to content. Since $P_v CO_2$ and PBF remain unchanged with this maneuver, these equations together yield a single equation:

$$\text{Eq. 4: } PBF = (\dot{V}_{CO_2} - \dot{V}_{CO_2}') / (S \times P_A CO_2' - S \times P_A CO_2)$$

This way, a small, temporary change in \dot{V}_A can be used to calculate PBF without the need for invasive monitors. This method has been proven reliable in intubated patients by multiple studies [10-12].

Creative as this method is; however, there are several limitations to it. First, the time to reach the new steady state is a function of the subject's V_A and functional residual capacity (FRC). On average, this time is approximately 30 - 45 s, or about five breaths, which means that only two data points are used to calculate PBF before recirculation of arterial blood occurs, limiting the method's accuracy. Second, because V_A is required to be transiently reduced, and because the normal variability of breathing creates noise in the measurement signal, this method is limited only to intubated patients. Finally, $P_A CO_2$ must be allowed to re-equilibrate to steady state levels before a subsequent test can be performed, which may require an additional 60 - 120 s.

Materials and Methods

Flux of CO₂ at the alveoli

The method proposed here redefines the term \dot{V}_{CO_2} as described in Fick's equation. According to conventional understanding, \dot{V}_{CO_2} is defined as the excretion of CO₂ past the lips, measured with a pneumotachometer and CO₂ monitor. However, exhaled CO₂ on a breath-by-breath basis seldom reflects actual metabolic CO₂ production or pulmonary capillary excretion into the lungs [13]. Instead, an average over several breaths is required for this estimation, as the FRC acts as a reservoir for CO₂. For example, a large breath that expunges a large volume of CO₂ past the lips would incorrectly reflect a large "production" of CO₂. The Fick equation, therefore,

Table 1. The Parameters Used in the Mathematical Model

Parameter	Error	Value
PvCO ₂ (mm Hg)	None	50
Pulmonary blood flow (L/min)	None	6
Tidal volume (mL)	±30%	500
Respiratory frequency (/min)	None	12
FRC (L)	None	3
PBF-preset (L/min)	None	6

would be more physiologically accurate if the term VCO₂ represented the continuous flux of CO₂ from the blood into the alveoli (VCO_{2A}), rather than the discrete, tidal excretion of CO₂ past the lips (VCO_{2M}). Averaged over many breaths, VCO_{2M} will accurately reflect VCO_{2A}.

Relationship between VCO_{2A} and P_ACO₂

After redefining VCO₂ in this manner, it becomes clear that there exists a linear relationship between the flux of CO₂ out of the blood (VCO_{2A}) and P_ACO₂ (Fig. 1).

At steady state VCO_{2A} and P_ACO₂ produce a single point on this line. The slope of this line, PBF, is proportional to the PBF. A second theoretical line exists on this diagram, connecting the origin with the steady state point. This second line is described by the basic physiologic equation:

$$\text{Eq. 5: } VCO_{2A} = V_A \times P_A CO_2 / 713$$

where V_A is the alveolar ventilation - the slope of the line. It becomes clear that VCO_{2A} - not VCO_{2M} - must be applied when describing these relationships, as washout of the FRC can confound the changes that relate these variables.

Gedeon's method aims to use two steady state VCO_{2M} points to determine the slope of the PBF line: one attained from the subject breathing at rest (where VCO_{2M} only equals VCO_{2A} if averaged over many breaths), and the second after a small change in V_A (measured after 4 - 5 breaths, when, after a transient period, VCO_{2M} is assumed to be equal to VCO_{2A}). In an extreme case, if the breath were held for 30 s and a new steady state were reached, P_ACO₂ would be equal to PvCO₂, or the x-intercept (Fig. 2). This method can only measure two points because after 4 - 5 breaths of breathing at a second V_A, recirculation would alter PvCO₂. However, if VCO_{2A} can be measured with each breath, one data point along this line could be produced with each exhalation, and the slope, PBF could be calculated with sequential breaths. Variations in VCO_{2A} are needed to explore this line, and the slope of the second line, V_A, would vary with each breath due to normal variations in tidal volume and respiratory frequency. Since VCO_{2A} likely fluctuates little during quiet breathing (as opposed to VCO_{2M}), variability of breathing, which was once the system "noise", therefore becomes important for measurement of PBF.

Once this line is observed over multiple breaths, PBF can be calculated by extrapolating to the x-intercept, PvCO₂, and calculating PBF using equation 1.

Method for calculating VCO_{2A}

The method described above requires calculation of the VCO_{2A}, the flux of CO₂ across the alveolar membrane. Calculation for VCO_{2A} has been described thoroughly elsewhere [13]. Briefly, a simple mass balance dictates that the flux of CO₂ into the alveoli from the blood is related to the flux of CO₂ at the mouth and the change of CO₂ stores in the lung:

$$\text{Eq. 6: } VCO_{2A} = VCO_{2M} - \Delta VCO_{2S}$$

where VCO_{2A} is the alveolar flux of CO₂, VCO_{2M} is the flux of CO₂ at the mouth, and ΔVCO_{2S} is the change in lung stores of CO₂. VCO_{2M} is easily measured using a metabolic cart, which integrates exhaled CO₂ concentration and expiratory flow.

The change in lung stores of CO₂ from breath 1 to breath 2, ΔVCO_{2S}, can be described as the sum of two terms describing the change in CO₂ concentration at a constant V_A, and the change in V_A at a constant CO₂ concentration:

$$\text{Eq. 7: } \Delta VCO_{2S} = V_A (P_A CO_2' - P_A CO_2) / 713 + P_A CO_2' \Delta V_A$$

where ΔVCO_{2S} is the change in alveolar stores of CO₂, V_{A1} is the alveolar volume at the end of breath 1, P_ACO₂ and P_ACO₂' are the fractions of CO₂ in the alveoli at the end of breaths 1 and 2, respectively and ΔV_A is the change in alveolar volume from breath 1 to breath 2.

Computer simulation

For this study, a computer simulation of tidal breathing was designed to test the theory under ideal conditions using Microsoft Excel 2003 (Redmond, Washington, USA). Incremental calculations of lung CO₂ volume were made for each 0.001 min. Tidal breathing was simulated using variable inhaled and exhaled tidal volumes (±30%), allowing lung volume to return to a different FRC with each breath. Complete alveolar mixing was assumed, inspiratory and expiratory times were equal, inhalation and exhalation flows were linear, and alveolar dead-space and shunt were assumed to be minimal. P_ACO₂ was recorded once per breath (the final P_ACO₂ value at the end of exhalation) and VCO_{2M} was calculated at the mouth during exhalation. The model was run over a period of 10 min (100 breaths), and P_ACO₂, VCO_{2M}, Vt-in and Vt-out were recorded with each breath. V_A for sequential breaths was calculated as V_A' = V_A + Vt-in - Vt-out, where Vt-in and Vt-out are inhaled tidal volume and exhaled tidal volume, respectively. Detailed parameters are outlined in Table 1.

P_ACO₂-ave versus P_{ET}CO₂

As stated above, the diffusion of CO₂ across the alveolar membrane should form a linear relationship with the pressure of CO₂ in the alveoli. Therefore, measurements of P_{ET}CO₂, a single sample taken at the end of exhalation, would be inappropriate to use in equation 3 since it reflects an end-expiration value rather than an average (P_ACO₂-ave). In the computer model, an average P_ACO₂ is easy to calculate, but this is not the case clinically, when P_{ET}CO₂ samples are the only measurements




Figure 3. Alveolar partial pressure of CO₂ over a period of 2 min as predicted by the computer model.

readily available non-invasively. To estimate $P_A\text{CO}_2\text{-ave}$ for a single breath using only non-invasive data, we back-calculated to the $P_A\text{CO}_2$ that might exist at peak inhalation and averaged this with $P_{\text{ET}}\text{CO}_2$. The $P_A\text{CO}_2$ at peak inhalation might be estimated as:

Eq. 8: $(\text{Volume of CO}_2 \text{ in Lung at End-Exhalation} - \text{Volume of CO}_2 \text{ Diffused during Exhalation} + \text{Exhaled Volume of CO}_2) \times 713 / \text{Peak Alveolar Volume}$

Averaging this value with $P_{\text{ET}}\text{CO}_2$ may produce a reasonable estimate of $P_A\text{CO}_2\text{-ave}$ which can be applied in equation 1.

Calculation of VCO_{2A}

VCO_{2S} was calculated using equation 6 using sequential breaths, ΔV_A was assumed to be measurable without the need for nitrogen monitors ($V_t\text{-in}$ and $V_t\text{-out}$ were measurable), and V_A was initially assumed to be 3 L. VCO_{2A} was then calculated using equation 5. Using 10 min of breathing data, when PBF and $P_v\text{CO}_2$ were assumed to be in steady state, a plot of VCO_{2A} versus $P_A\text{CO}_2$ was created and R^2 was calculated for




Figure 4. VCO_2 over time as measured at the mouth (VCO_{2M} , open circles), as pre-set by the model ($\text{VCO}_{2A}\text{-model}$, open squares) and as calculated using the proposed method ($\text{VCO}_{2A}\text{-calculated}$, closed circles).




Figure 5. The relationship between VCO_{2A} and $P_A CO_2$ using end-tidal PCO_2 ($P_{ET} CO_2$, gray squares) versus time averaged CO_2 across the breath ($P_A CO_2$ -ave, closed circles).

the line of best fit. FRC was then iterated in increments of 0.25 L from 2 L to 4 L to achieve the most optimal linear relationship between VCO_{2A} and $P_A CO_2$.

Real-world adjustments

To simulate real-world measurements, error was introduced into each measurement ($P_A CO_2$, V_t -in, V_t -out, VCO_{2M} , Table 1), consistent with manufacturer specifications [14] and calculation of PBF was repeated.

PBF comparisons

PBF was calculated from equation 3 using two techniques: first, using known VCO_{2A} measurements directly from the model (PBF-preset), calculated using actual diffusion of CO_2 across the alveolar membrane, and second, using VCO_{2M} to calculate VCO_{2A} using equation 5 (PBF-calc), which represents the strategy that might be used practically in subjects. In the either case, PBF would be calculated using equation 1 where $P_v CO_2$ was attained by extrapolating to the x-intercept of a plot relating VCO_{2A} and $P_A CO_2$ -ave. We determined the appropriate number of breaths required for accurate measurement of PBF-calc, by increasing the number of breaths used to calculate the average PBF-calc until the repeatability was similar to that of thermodilution [15].

PBF-calc was assessed and evaluated using Bland-Altman analysis where the acceptable error was taken from prior established theory [16].

Results

The model for tidal ventilation produces reasonable values for $P_A CO_2$ over 10 min of breathing. Sample oscillations in $P_A CO_2$ and lung volume can be seen in Figure 3.

VCO_{2A} tended to vary less over the course of 10 min of breathing than VCO_{2M} , as can be seen in Figure 4. When VCO_{2A} was plotted against $P_A CO_2$ -ave a linear relationship was revealed which was stronger than with $P_{ET} CO_2$ ($R = 0.83$ versus 0.74, Fig. 5).

Iterating FRC demonstrated that small deviations from the model value produced poorer relationships between VCO_{2A} and $P_A CO_2$ (Fig. 6). However, the optimal FRC achieved after iterating FRC was 3.25 L, which was 0.25 L (8.3%) greater than the actual FRC used in the model. Non-optimal FRCs did not significantly reduce the accuracy of the calculated PBF, but did increase its variability.

When $P_A CO_2$ -ave was used instead of $P_{ET} CO_2$, the relationship was improved ($R^2 = 0.83$ versus 0.74). PBF-calc was calculated using 10 breaths at a time (see repeatability below) by extrapolating the line in Figure 3 to the x-intercept ($P_v CO_2$) and applied into equation 1. Using this method, the difference between PBF-preset and PBF-calc was 0.62 ± 0.53 L/min.

Figure 7 shows that the repeatability coefficient fell from 1.13 to 0.09 L/min as the number of breaths used to calculate PBF-calc increased from 2 to 30. The repeatability coefficient decreased and leveled at 0.09 when approximately 10 breaths were used to average calculation for PBF-calc. The repeatability coefficient for thermodilution is likely between 0.4 and 0.6 L/min [4, 15]), which is the equivalent of using 5 - 6 breaths to average PBF-calc measurements in the proposed method.




Figure 6. The R^2 statistical parameter relating VCO_{2A} to P_ACO_2 varies depending on the original estimate of FRC. As the initial guess of FRC is increased, R^2 reaches a peak, optimum point and then falls.

Discussion

The most common method for cardiac output measurement involves indicator dilution, normally dyes and temperature. A dye, or cold saline, is typically injected via an invasive catheter into the pulmonary artery and the temporal profile of the concentration of the dye or the temperature of the blood is measured downstream. The profile of the indicator change over time is used to calculate the cardiac output. Over the last three decades, there has been a steady improvement in the technology required to manufacture the catheters and thermistors, and to analyze the indicator curves and calculate the cardiac output [17]. In addition, the expertise to place the catheters and look after catheterized patients has become widespread.

Nevertheless, there are at least three drawbacks necessarily associated with these methods. First, pulmonary artery catheters are inherently invasive and have associated complications including damage to the carotid artery, subclavian artery and lung, air emboli, pneumothorax, malignant arrhythmias and heart block, rupture of right atrium, right ventricle and/or pulmonary capillary, local infection and septicaemia, and more [3]. Second, they have many associated costs as a result of the requirement for critical care areas, equipment and personnel. Third, their accuracy is questionable, and can be less reliable and helpful as required for management of critically ill patients [18]. Despite these drawbacks, thermodilution methods continue to be widely used as less invasive alternatives are not sufficiently accurate, not sufficiently robust, or too cumbersome to perform in a large variety of clinical settings [19].

In this study, a new non-invasive method of measuring

PBF is introduced based on principles that are already accepted in medicine. The Fick method for measuring PBF is well established and is considered one of the most accurate techniques available. Until recently, however, the Fick technique could only be performed using blood samples despite numerous attempts to measure $PvCO_2$ non-invasively. The method of creating a step-change in V_A is the only validated non-invasive Fick method of cardiac output measurement without a special maneuver required by the patient, but can only be used on patients ventilated by a mechanical ventilator, uses only two breaths for measurement ETC. If a spontaneously breathing patient is made to rebreathe previously exhaled gas, the minute ventilation will tend to increase in order to increase the volume of air entering the lungs for gas exchange. The method presented here would require no maneuver on the part of the subject, and no foreign or compressed gases.

The method outlined here describes an original relationship between VCO_2 and P_ACO_2 . Its principles are grounded in basic physiology, and its application may extend beyond that of PBF measurement. Since P_ACO_2 relates directly to VCO_2 given a specific set of conditions, other variables that may influence these parameters such as alveolar dead space, may be measurable as well.

There are several practical limitations of this method. First, this strategy for measuring PBF could not be achieved without a perfect, air-tight seal around the mouth, nose, or face. Any air lost would reduce the accuracy of Vt -in, Vt -out, or VCO_{2M} . This may be inconvenient or impossible for some patients, depending on the presence of facial hair, anomalous anatomy, or trauma.




Figure 7. As few as two breaths can be used to calculate PBF but its reliability is increased as more breaths are incorporated into the calculation. Precision reaches a plateau at about 8 - 10 breaths.

This model assumes that all exhaled gas had participated in exchange of CO_2 with the blood. P_ACO_2 may vary depending on differences in ventilation-to-perfusion matching throughout the lung [2, 20]. If some exhaled gas originated from alveolar dead space, $\text{P}_{\text{ET}}\text{CO}_2$ would underestimate P_ACO_2 and as a result, PBF-calc would overestimate PBF. The significance of this was not quantified in this study, but theoretically, it is possible that iterations of alveolar dead space estimates can be coupled with the estimates of FRC to provide an optimal relationship between P_ACO_2 and VCO_{2A} . Still, deviations of $\text{P}_{\text{ET}}\text{CO}_2$ from P_ACO_2 may also be due to incomplete breaths during exhalation and not alveolar dead space per se. This also impacts the calculation for $\text{P}_A\text{CO}_{2\text{-ave}}$, which will also be affected by all of the above.

In this study, 10 min of breathing were permitted to achieve initial FRC estimate. At this point, it is unclear why the ideal FRC produced from the iterative process differed from that used in the model. Nevertheless, its impact on PBF-calc was marginal. Furthermore, the purpose of this extended period of breathing was to demonstrate the principle rather than practicality. Follow-up studies in humans will be required to further explore this strategy.

The real-world error in measuring Vt , $\text{P}_{\text{ET}}\text{CO}_2$, VCO_2 was based on products currently available for purchase. These will vary depending on the manufacturer and may improve in the future, making this method more practical.

The accuracy of this method for measuring PBF was similar to others that have been proposed [15]. The repeatability demonstrated that 10 breaths were needed for optimal accuracy, a time slightly greater than would be required for a complete test of thermodilution (about 30 s) depending on respiratory frequency. However, one advantage of the proposed method in this respect is that measurements for PBF are continuous, and no time is needed for “reset”, as may be needed to washout indicator from the pulmonary artery for thermodilution.

Finally, whereas the accuracy of other respiratory-based non-invasive methods may be diminished by respiratory fluctuations, the proposed method is enhanced by large changes in tidal volume and respiratory frequency. It may even be suggested that a subject ought to be encouraged to take deep breaths, or sigh to exaggerate the variability of quiet tidal breathing.

Conclusion

The study we describe here is safe, theoretically sound, and demonstrates acceptable accuracy and repeatability. It represents a potential advancement towards measurement of important cardiopulmonary parameters that may be clinically important in the management of both outpatients and those in-hospital. Further studies in humans are required to quantify and evaluate its strengths and limitations.

Acknowledgement

None.

Conflict of Interest

The authors report no conflicts of interest.

Source(s) of Support

Funding for this study was provided by internal department

funds.

References

1. Branthwaite MA, Bradley RD. Measurement of cardiac output by thermal dilution in man. *J Appl Physiol.* 1968;24(3):434-438.
2. Kim TS, Rahn H, Farhi LE. Estimation of True Venous and Arterial Pco₂ and Cardiac Output by Gas Analysis of a Single Breath. *Techn Docum Rep Amrl-Tdr-63-103 (I).* AMRL TR. 1963:58-73.
3. Nadeau S, Noble WH. Limitations of cardiac output measurements by thermodilution. *Can Anaesth Soc J.* 1986;33(6):780-784.
4. Wong LS, Yong BH, Young KK, Lau LS, Cheng KL, Man JS, Irwin MG. Comparison of the USCOM ultrasound cardiac output monitor with pulmonary artery catheter thermodilution in patients undergoing liver transplantation. *Liver Transpl.* 2008;14(7):1038-1043.
5. Wendelboe Nielsen O, Hansen S, Christensen P, Gronlund J. Repeatability of the acetylene rebreathing method in measuring cardiac output: influence of acetylene concentration. *Acta Anaesthesiol Scand.* 1990;34(5):354-357.
6. Fick A. Ueber die Messung des Blutquantums in den Herzventrikeln. *Sitzungsberichte der Physiologisch Medizinosche Gesellschaft zu Wuerzburg.* 2.
7. Levett JM, Replogle RL. Thermodilution cardiac output: a critical analysis and review of the literature. *J Surg Res.* 1979;27(6):392-404.
8. Olszowka AJ, Farhi LE. A system of digital computer subroutines for blood gas calculations. *Respir Physiol.* 1968;4(2):270-280.
9. Gedeon A, Forslund L, Hedenstierna G, Romano E. A new method for noninvasive bedside determination of pulmonary blood flow. *Med Biol Eng Comput.* 1980;18(4):411-418.
10. Bosman RJ, Stoutenbeek CP, Zandstra DF. Non-invasive pulmonary blood flow measurement by means of CO₂ analysis of expiratory gases. *Intensive Care Med.* 1991;17(2):98-102.
11. Capek JM, Roy RJ. Noninvasive measurement of cardiac output using partial CO₂ rebreathing. *IEEE Trans Biomed Eng.* 1988;35(9):653-661.
12. Peyton PJ, Thompson D, Junor P. Non-invasive automated measurement of cardiac output during stable cardiac surgery using a fully integrated differential CO₂ Fick method. *J Clin Monit Comput.* 2008;22(4):285-292.
13. Auchincloss JH, Jr., Gilbert R, Baule GH. Effect of ventilation on oxygen transfer during early exercise. *J Appl Physiol.* 1966;21(3):810-818.
14. CareFusion I. Manufacturer Specifications. Vmax System Specifications. http://www.carefusion.com/pdf/Respiratory/Pulmonary_Function_Testing/RC1355-01_Vmax-Specsheet10.pdf. 2013.
15. Critchley LA, Critchley JA. A meta-analysis of studies using bias and precision statistics to compare cardiac output measurement techniques. *J Clin Monit Comput.* 1999;15(2):85-91.
16. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet.* 1986;1(8476):307-310.
17. Lenfant C. Time-dependent variations of pulmonary gas exchange in normal man at rest. *J Appl Physiol.* 1967;22(4):675-684.
18. Nishikawa T, Dohi S. Errors in the measurement of cardiac output by thermodilution. *Can J Anaesth.* 1993;40(2):142-153.
19. Stetz CW, Miller RG, Kelly GE, Raffin TA. Reliability of the thermodilution method in the determination of cardiac output in clinical practice. *Am Rev Respir Dis.* 1982;126(6):1001-1004.
20. Dubois AB. Alveolar CO₂ and O₂ during breath holding, expiration, and inspiration. *J Appl Physiol.* 1952;5(1):1-12.